QuickFlux: Hookup and Startup

This guide walks you through the step-by-step process of setting up and starting your QuickFlux system, from connecting the batteries and hose to powering on the equipment. Following these instructions will ensure your system is properly connected, warmed up, and ready for animal measurements.

WATCH VIDEO

Steps

1. Attach the first battery. Make sure it clicks into place.
2. Attach the second battery. Make sure it clicks into place.
3. Fit the hose fully over the attachment on the bottom of the unit.
4. Use the quick clamp to lock the hose in place.
5. Press the green button on the RFID reader to turn it on.
6. If an ID scan starts on its own, ignore or cancel it and go back to the reader screen.
7. Turn on the QuickFlux unit. (Do this only after the batteries, hose and reader are ready.)
8. The light turns solid red while the unit warms up.
9. Wait about 10 minutes before bringing animals into the testing area.
10. When warm-up is done, the light blinks blue. The unit is now ready.